

NUS CORPORATION  
SUPERFUND DIVISION

FILE COPY  
CONFIDENTIAL

07571

INTERNAL CORRESPONDENCE

TO: FILE (CONFIDENTIAL)  
FROM: JOANNE MORIN JOM  
SUBJECT: OVA SCREEN-INTERSTATE UNIFORM  
TDD No. F1-8303-07A  
Job No. 3425 03 06  
0300.01

DATE: DECEMBER 28, 1983

COPIES FILE

Superfund Records Center

SITE: Wells GH

BREAK: 3.1

OTHER: 587742

On December 19, 1983, I screened three samples collected from Interstate Uniform Well No. 1 on December 8, 1983.

Sample No.

Sample Description

53497

ERT Field Blank

53440

Second bailfull

53443

Groundwater sample

The following Organic Vapor Analyzer (OVA) GC conditions were used: G-24 column, room temperature and injection size -1000 ul. No chromatographic peaks were detected. Notably, no peaks were recorded at the retention time of tetrachloroethylene. The chromatograms are attached.

JOM/mth  
Attachment



